

PROFICIENCY TEST « RAEMA »



réseau d'analyses et d'échanges en microbiologie des aliments

SCHEME N° 73 (4th OCTOBER 2021) GENERAL REPORT

« Any reproduction of the report must be made in its entirety»
« The Cofrac logo may not be used outside this report»

V. CARLIER⁽¹⁾, L. ALI-MANDJEE et M. CARLIER
ASA - ENVA, 7 avenue du Général de Gaulle, 94704 MAISONS ALFORT CEDEX

⁽¹⁾ Coordinator of the proficiency test « RAEMA »

For any claim, you can use the specific form available on our Website
<https://association.asa-spv.fr>

Table of contents

1- GENERAL DATA.....	3
1-1 PARTICIPATING LABORATORIES	3
1-2 DELIVERY TIME OF THE PARCEL	3
1-3 INFORMATION ABOUT SAMPLE	3
1-3-1 NATURE	3
1-3-2 SIZE	3
1-3-3 HOMOGENEITY AND STABILITY OF THE CONTAMINATION.....	3
1-3-4 FLORA FOR ENUMERATION / DETECTION.....	3
1-4 EXECUTION OF ANALYSIS	4
1-4-1 DELIVERY TIME OF SAMPLES / BEGINNING OF ANALYZES	4
1-4-2 PRESERVATION TEMPERATURE OF SAMPLE BEFORE ANALYSIS	4
2- EXPLOITATION OF ANALYSIS REPORT.....	4
2-1 PREPARATION OF THE INITIAL SUSPENSION.....	4
2-2 DILUENT USED FOR THE INITIAL SUSPENSION	4
2-3 HOMOGENIZATION TECHNIQUE	4
2-4 RESUSCITATION'S CONDITIONS	4
2-4-1 DURATION	4
2-4-2 TEMPERATURE.....	4
2-5 MICROORGANISMS AT 30°C	5
2-6 ENTEROBACTERIACEA	6
2-7 TOTAL COLIFORMS	7
2-8 THERMOTOLERANT COLIFORMS.....	8
2-9 ESCHERICHIA COLI.....	9
2-10 ANAEROBIC SULFITE-REDUCING BACTERIA.....	10
2-11 CLOSTRIDIUM PERFRINGENS	11
2-12 COAGULASE POSITIVE STAPHYLOCOCCI	12
2-13 LISTERIA MONOCYTOGENES – ENUMERATION	13
2-14 SALMONELLA –DETECTION.....	15
2-15 LISTERIA MONOCYTOGENES –DETECTION	17
3- ASSESSMENT OF PERFORMANCE	19
3-1 PERFORMANCES IN ENUMERATION.....	19
3-1-1 MICROORGANISMS AT 30°C	21
3-1-2 ENTEROBACTERIACEA	21
3-1-3 TOTAL COLIFORMS.....	21
3-1-4 THERMOTOLERANT COLIFORMS	22
3-1-5 ESCHERICHIA COLI	22
3-1-6 ANAEROBIC SULFITE-REDUCING BACTERIA.....	22
3-1-7 CLOSTRIDIUM PERFRINGENS	23
3-1-8 COAGULASE POSITIVE STAPHYLOCOCCI.....	23
3-1-9 LISTERIA MONOCYTOGENES – ENUMERATION	23
3-2 PERFORMANCES IN DETECTION.....	24
3-2-1 DETECTION – SALMONELLA	24
3-2-2 DETECTION - LISTERIA MONOCYTOGENES.....	24
3-3 EVOLUTION OF PERFORMANCE	24

1-GENERAL DATA

1.1.PARTICIPATING LABORATORIES

345 laboratories participated to the 73th scheme. The sending was made on Monday 4th October 2021. We received **338** answers (98.0%).

1.2.DELIVERY TIME OF THE PARCEL

Reception	J0	J0+1	J0+2	J0+3	J0+4	J0+5	J0+6	J0+7	J0+8	J0+9	J0+11	J0+18
Nb of laboratories	6	235	34	21	16	4	1	7	5	7	1	1

1.3.INFORMATIONS ABOUT SAMPLE

1.3.1. NATURE

The sample included :

- one strain of *Enterococcus sp.* at a concentration level of 10^5 cfu/g in 5 units ;
- one strain of *Citrobacter sp.* at a concentration level of 5.10^2 cfu/g in 5 units ;
- one strain of *Serratia liquefaciens* at a concentration level of 5.10^2 cfu/g in 5 units ;
- one strain of *Escherichia coli* at a concentration level of $1.5.10^2$ cfu/g in 5 units ;
- one strain of *Clostridium perfringens* at a concentration level of 5.10^2 cfu/g in 3 units ;
- one strain of *Staphylococcus aureus* at a concentration level of 3.10^3 cfu/g in 5 units ;
- one strain of *Salmonella Anatum* at a concentration level of 25 cfu/g in 2 units ;
- one strain of *Listeria monocytogenes* at a concentration level of 2.10^3 cfu/g in 3 units .

1.3.2. SIZE

180 kilogrammes of powder were produced, and distributed after contamination in bottles containing 75 grammes at least. Bottles were covered by a label with a 6 digit identification number.

1.3.3. HOMOGENEITY AND STABILITY TEST OF THE CONTAMINATION

Homogeneity and stability of samples are checked during the statistical analysis of participants results. A supplementary check of the contamination's homogeneity was realized on 10 samples for each unit by a double enumeration of aerobic microorganisms at 30°C.

The contamination's stability was also checked by enumeration / detection of all flora on 11, 18 and 25 October 2021. These checks were realized by a subcontractor accredited by Cofrac.

Homogeneity and stability of samples have been validated.

1.3.4. FLORA FOR ENUMERATION / DETECTION

Enumeration of the following flora was proposed : microorganisms at 30°C, Enterobacteriaceae, total and thermotolerant coliforms, beta-glucuronidase positive *Escherichia coli*, anaerobic sulfite-reducing bacteria, *Clostridium perfringens*, coagulase positive staphylococci, *Listeria monocytogenes*, as well as detection of *Salmonella* and *Listeria monocytogenes*.

1.4. EXECUTION OF ANALYZES

1.4.1. DELIVERY TIME OF SAMPLES / BEGINNING OF ANALYZES

338 laboratories (100%) specified it.

Analysis time	J0	J0+1	J0+2	J0+3	J0+4	J0+5	J0+6	J0+7	J0+8	J0+9	J0+10	J0+11	J0+14	J0+15	J0+18	J0+21
Nb of laboratories	1	41	42	32	4	1	2	135	46	14	7	3	5	3	1	1

1.4.2. PRESERVATION TEMPERATURE OF SAMPLES BEFORE ANALYSIS

333 laboratories (98.5%) specified it. The average temperature is **3.9°C** with a standard deviation of 0.9°C. The given data 19, 20, 21, 22, 25, 26 and 26.4°C given by 10 laboratories were not taken into account for this calculation.

2. EXPLOITATION OF ANALYSIS REPORT

2.1. PREPARATION OF THE INITIAL SUSPENSION

For **338** answers (100%) :

209 laboratories (61.8%) prepare the initial suspension with adding diluent to powder.

128 laboratories (37.9%) prepare the initial suspension with adding powder to diluent.

1 laboratory (0.3%) used another technique to prepare the initial suspension.

2.2. DILUENT USED FOR THE INITIAL SUSPENSION

For **335** answers (99.1%) :

281 laboratories (83.1%) use Buffered Peptone Water for the initial suspension.

40 laboratories (11.8%) use Peptone salt for the initial suspension.

14 laboratories (4.1%) used another diluent for the initial suspension. Among these 14 laboratories, 9 specified utilization of Salmonella enrichment broth.

2.3. HOMOGENEIZATION TECHNIQUE

For **336** answers (99.4%) :

313 laboratories (92.6%) homogenize their sampling with a StomacherND.

23 laboratories (6.8%) used another technique (manual, magnétic or other).

2.4. RESUSCITATION'S CONDITIONS

2.4.1. DURATION

324 laboratories (95.9%) specified it.

The average duration is **26.7 min** with a standard deviation of 15.5 min. The data 90, 120 and 1440 min given by 3 laboratories were not taken into account for this calculation.

2.4.2. TEMPERATURE

324 laboratories (95.9%) specified it.

The average temperature is **21.3°C** with a standard deviation of 2.7°C.

2.5. MICROORGANISM AT 30°C

321 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
	ISO/NF EN ISO 4833-1	197
	AFNOR 3M-01/1-09/89	50
	NM ISO 4833-1	23
	ISO/NF EN ISO 4833-2	16
	AFNOR BIO-12/35-05/13	12
	XP V08-034	6
	Other	17
	+ V08-100 (spiral)	20
Culture medium		
	Plate Count Agar	233
	Petrifilms	50
	Plate Count Agar + Milk	23
	Tempo AC	12
	Other	1
Preparation		
	Home made	108
	Ready to use not pre-poured	140
	Ready to use, plate, film, card	71
Plating method		
	Surface	70
	Pour	230
	Culture medium for card	13
1st dilution retained		
	- 1	18
	- 2	17
	- 3	262
	- 4	12
	1/400	6
	1/4000	4
Incubation temperature		
	30°C	315
	33-35°C	2
	37°C	1
	25°C	1
Incubation duration		
	68-75 h	263
	44-48 h	48
	24-26 h	5
	40 h	2
	120 h	1

2.6.ENTEROBACTERIACEA

279 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
NF V08-054		107
→ NM 08.0.109 ⁽¹⁾		20
ISO/NF EN ISO 21528-2		73
AFNOR 3M-01/6-09/97		46
AFNOR BIO-12/21-12/06		11
AFNOR AES-10/07-01/08		8
AFNOR BRD-07/24-11/13		8
Other		6
	+ V08-100 (spiral)	4
Culture medium		
VRBG		200
Petrifilms		49
Tempo EB		10
Rebecca		10
Rapid'Enterobacteriaceae		8
Other		2
Preparation		
Home made		83
Ready to use not pre-poured		137
Ready to use, plate, film, card		59
1st dilution retained		
- 1		212
- 2		59
1/40		3
1/400		5
Incubation temperature		
37°C		178
30°C		89
35°C		11
Incubation duration		
20-26 h		272
48 h		6
Confirmatory test		
Yes		66
No		208

⁽¹⁾ Similar method to NF V08-054 according to ONSSA (Office National de Sécurité Sanitaire des produits Alimentaires).

2.7.TOTAL COLIFORMS

237 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
	NF V08-050	109
	→ NM 08.0.142 ⁽²⁾	9
	ISO/NF ISO 4832	59
	AFNOR 3M	24
	NM ISO 4832	20
	AFNOR BIO-12/17-12/05	6
	AFNOR BRD-07/08-12/04	4
	Other	6
	+ V08-100 (spiral)	2
Culture medium		
	VRBL	199
	Petrifilms	25
	Tempo TC	6
	Rapid Ecoli	5
	Other	2
Preparation		
	Home made	84
	Ready to use not pre-poured	124
	Ready to use, plate, film, card	29
1st dilution retained		
	-1	201
	-2	30
	1/40	2
	1/400	2
Incubation temperature		
	30°C	219
	37°C	17
Incubation duration		
	18.1-27 h	233
	48 h	3

AFNOR 3M method including :

5 laboratories specified utilization of AFNOR 3M-01/02-09/89 A method.

1 laboratory specified utilization of AFNOR 3M Petrifilm CC method.

⁽²⁾ Similar method to NF V 08-050 according to ONSSA.

2.8.THERMOTOLERANT COLIFORMS

219 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method	NF V08-060 → NM 08.0.124 ⁽³⁾ AFNOR 3M ISO/NF ISO 4832 Other	146 27 28 14 4
	+ V08-100 (spiral)	3
Culture medium	VRBL Petrifilms Other	188 29 2
Preparation	Home made Ready to use not pre-poured Ready to use, plate, film, card	79 114 26
1st dilution retained	-1 -2	195 23
Incubation temperature	42-45°C 37°C	216 2
Incubation duration	21-24 h 48 h 34 h	214 3 1

AFNOR 3M method including :

- 6 laboratories specified utilization of AFNOR 3M-01/02-09/89 C method.
- 1 laboratory specified utilization of AFNOR 3M-01/05-03/97 B method.
- 1 laboratory specified utilization of AFNOR 3M Petrifilm CC method.
- 1 laboratory specified utilization of AFNOR 3M Petrifilm high sensitivity method.
- 1 laboratory specified utilization of AFNOR 3M Petrifilm EC method.

⁽³⁾ Similar method to NF V08-060 according to ONSSA.

2.9.ESCHERICHIA COLI

295 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
	ISO/NF ISO 16649-2	170
	AFNOR 3M	41
	NM ISO 16649-2	22
	AFNOR BRD-07/01-07/93	18
	AFNOR BIO-12/13-02/05	11
	AFNOR AES-10/06-01/08	9
	NM 08.0.108	6
	AFNOR BIO-12/05-01/99	5
	NF EN ISO 16649-3	1
	Other	10
	+ V08-100 (spiral)	4
Culture medium		
	TBX	200
	Petrifilms	42
	Rapid E. coli	21
	Rebecca	12
	Tempo EC	11
	Coli ID	7
	Other	1
Preparation		
	Home made	84
	Ready to use not pre-poured	159
	Ready to use, plate, film, card	50
Plating method		
	Surface	43
	Pour	232
	Culture medium for card	12
1st dilution retained		
	-1	275
	-2	9
	1/40	3
	1/400	6
Incubation temperature		
	41-46°C	257
	37°C	34
	30°C	1
Incubation duration		
	18-26 h	288
	48 h	4

AFNOR 3M method including :

- 17 laboratories specified utilization of AFNOR 3M-01/08-06/01 (*SELECT'E. COLI*) method.
- 1 laboratory specified utilization of AFNOR 3M-01/04-09/92 method.
- 1 laboratory specified utilization of AFNOR 3M Petrifilm EC method.

2.10. ANAEROBIC SULFITE-REDUCING BACTERIA

241 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
	NF V08-061	158
	→ NM 08.0.125 ⁽⁴⁾	15
	ISO/NF ISO 15213	41
	NM ISO 15213	13
	Other	14
Culture medium		
	TSC	220
	TSN	10
	Iron Sulfite agar	7
	Other	3
Preparation		
	Home made	91
	Ready to use not pre-poured	123
	Ready to use, plate, film, card	27
Seeding way		
	Plates	150
	Tubes	90
1st dilution retained		
	-1	159
	-2	75
	-3	3
Incubation temperature		
	44-50°C	173
	37±1°C	65
	30°C	3
Incubation duration		
	16-24 h	202
	43-48 h	31
	72 h	8

⁽⁴⁾ Similar method to NF V08-061 according to ONSSA.

2.11. CLOSTRIDIUM PERFRINGENS

192 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
ISO/NF EN ISO 7937		158
NM ISO 7937		24
NM 08.0.111		2
Other		8
Culture medium		
TSC		190
Other		1
Preparation		
Home made		65
Ready to use not pre-poured		122
Ready to use, plate, film, card		4
1st dilution retained		
-1		148
-2		41
-3		2
Incubation temperature		
37±1°C		181
44-46°C		9
30°C		1
Incubation duration		
17-27 h		184
48 h		7
Confirmation test		
None		30
Lactose-sulfite		132
Strip		9
Other		6

2.12. COAGULASE POSITIVE STAPHYLOCOCCI

298 laboratories performed the enumeration.

Parameters	Mode	Nb laboratories
Method		
	ISO/NF EN ISO 6888-2	139
	ISO/NF EN ISO 6888-1	67
	AFNOR BKR-23/10-12/15	21
	NM ISO 6888-1	20
	AFNOR 3M-01/9-04/03	17
	AFNOR BIO-12/28-04/10	11
	NM ISO 6888-2	5
	NM 08.0.112	3
	NordVal No :049	2
	Other	16
	+ V08-100 (spiral)	6
Culture medium		
	RPF	135
	BP+egg yolk tellurite	86
	Easy Staph	27
	Petrifilm	18
	BP+egg yolk tellurite+ sulfamethazine	13
	Tempo STA	11
	Rapid Staph	2
	Other	6
Preparation		
	Home made	65
	Ready to use not pre-poured	126
	Ready to use, plate, film, cards	104
Plating method		
	Surface	152
	Pour	132
	Culture medium for card	11
1st dilution retained		
	-1	119
	-2	164
	-3	3
	1/40	3
	1/400	5
Incubation temperature		
	37±1°C	294
	30°C	2
Incubation duration		
	40-48 h	194
	18-26 h	102
Confirmation test		
	None	185
	Staphylo-coagulase	78
	Clumping factor	16
	DNase	10
	Other	4

2.13. LISTERIA MONOCYTOGENES – ENUMERATION

244 laboratories performed the enumeration.

RESUSCITATION

78 laboratories announce the realization of a resuscitation step.

Details concerning temperature and average duration of this step are not required anymore in the input form.

Parameters	Mode	Nb laboratories
Method		
	ISO/NF EN ISO 11290-2	63
	AFNOR AES-10/05-09/06	59
	AFNOR BKR-23/05-12/07	48
	AFNOR BRD-07/05-09/01	29
	NM ISO 11290-2	27
	AFNOR BRD-07/17-01/09	10
	Other	8
Resuscitation medium		
	Buffered Peptone Water	198
	Fraser base	33
	Other	11
Enumeration medium		
	ALOA Count	114
	Compass Listeria	69
	Rapid Lmono	33
	AL Agar	16
	Palcam	6
	OCLA	4
	Other	2
Preparation		
	Home made	36
	Ready to use not pre-poured	49
	Ready to use, plate, film, card	158
Plating method		
	Surface	200
	Pour	40
	Culture medium for card	0

Parameters	Mode	Nb laboratories
1st dilution retained	-1	153
	-2	87
Incubation temperature	37±1°C	239
	30°C	3
	22°C	1
Incubation duration	42-48 h	193
	24±1 h	49
	1 h	1
Confirmation test	None	49
	Biochemical	134
	Biochemical + CAMP	39
	Other	11
Nb of colonies tested per plate	1	63
	2-3	11
	5	106

2.14. SALMONELLA – DETECTION

305 laboratories performed the detection.

Methods used by laboratories are clarified in the following table :

Parameter	Mode	Nb laboratories
Method		
	ISO/NF EN ISO 6579-1	78
	AFNOR BKR 23/07-10/11 (IRIS Salmonella)	72
	NM ISO 6579-1	31
	AFNOR BRD 07/11-12/05 (Rapid Salmonella)	29
	AFNOR BIO 12/32-10/11 (VIDAS SPT)	29
	AFNOR BIO 12/41-03/17 (SALMA One day)	27
	AFNOR BIO 12/16-09/05 (VIDAS Easy Salmonella)	20
	AFNOR UNI 03/07-11/13 (PCR)	2
	AFNOR BRD 07/06-07/04 (PCR)	1
	Other	16

No detail of methodology was asked to laboratories using other methods than ISO/NF EN ISO 6579-1 and NM ISO 6579-1 methods and proposed in the questionnaire.

You will find below a short description of these methods :

Method	Pre-enrichment	Enrichment	Isolation
AFNOR BIO 12/16-09/05 VIDAS Easy Salmonella	BPW / 37°C - 16/20h	SX2 / 41,5°C - 22/26h	Chrom ID / 37°C - 24h
AFNOR BIO 12/32-10/11 VIDAS SPT		BPW + Salmonella supplement / 41,5°C - 18/24h	Chrom ID / 37°C - 24h
AFNOR BKR 23/07-10/11 IRIS Salmonella		IRIS Salmonella Enrichment / 41,5°C - 18±2h	IRIS / 37°C - 24±3h
AFNOR BRD 07/11-12/05 Rapid Salmonella		BPW + Salmonella capsule / 41,5°C - 18±2h	Rapid Salmonella / 37°C - 24±2h
AFNOR BIO 12/41-03/17 SALMA One day		BPW + Salmonella supplement / 41,5°C - 16/24h	SALMA / 37°C - 24±3h
AFNOR UNI 03/07-11/13 PCR		BPW + supplement / 34-38°C – 20/24h	Lysis + PCR
AFNOR BRD 07/06-07/04 PCR		BPW / 37°C – 18/21h	Lysis + PCR

The detail of the methodology followed by 109 laboratories using ISO/NF EN ISO 6579-1 and NM ISO 6579-1 method, and the 16 laboratories using another method, is clarified in the following table :

Parameter	Mode	Nb laboratories
Method	ISO/NF EN ISO 6579-1	78
	NM ISO 6579-1	31
	Other	16
Pre-enrichment medium	None pre-enrichment	0
	Buffered Peptone Water	118
	Other	5
Pre-enrichment temperature	37±1°C	113
	41-42.5°C	9
	20-22°C	2
Pre-enrichment duration	16-20 h	88
	22-24 h	34
	30 h	1
	72 h	1
Enrichment medium	None enrichment	5
	RVS	107
	MKTn	105
	Selenite-cystine broth	21
	Other	7
Isolation medium	XLD	107
	Hektoen	33
	Bismuth Sulfate	23
	ASAP	18
	GVB	11
	SS	10
	IRIS Salmonella agar	9
	Rapid Salmonella	9
	Brilliance Salmonella	8
	Compass Salmonella	5
	Rambach	3
	Other	11
Confirmation test	Biochemical	45
	Biochemical + serological agglutination	70
	Other	7

2.15. LISTERIA MONOCYTOGENES – DETECTION

271 laboratories performed the detection.

Parameter	Mode	Nb laboratories
Method		
AFNOR AES 10/03-09/00 (ALOA one day)		65
ISO/NF EN ISO 11290-1		54
AFNOR BKR 23/02-11/02 (Compass L. mono)		49
AFNOR BRD 07/04-09/98 (Rapid' L. mono)		26
NM ISO 11290-1		23
AFNOR BIO 12/27-02/10 (VIDAS LMX)		10
AFNOR BIO 12/11-03/04 (VIDAS LMO2-37°C)		8
AFNOR BIO 12/02-06/94 (VIDAS Listeria)		7
AFNOR BRD 07/16-01/09 (Agar Listeria)		6
AFNOR BIO 12/18-03/06 (VIDAS LDUO)		4
AFNOR UNI 03/08-11/13 (PCR)		4
AFNOR BIO 12/40-11/16 (GENE UP LMO)		3
AFNOR UNI 03/04-04/05 (Listeria Precis)		2
Other		10

No detail of methodology was asked to laboratories using other method than ISO/NF EN ISO 11290-1 and NM ISO 11290-1 methods and proposed in the questionnaire.

You will find below a short description of these methods :

Méthod	Primary enrichment		Secondary enrichment		Isolation
	Medium	Incubation	Medium	Incubation	
AFNOR BRD 07/04-09/98 Rapid' L. mono	Fraser 1/2	30°C - 24±2h			Rapid L'mono 37°C - 24h
AFNOR BIO 12/02-06/94 VIDAS Listeria	Fraser 1/2	37°C - 26/30h	Fraser	30°C - 24/26h	Palcam et Oxford 37°C - 24h
AFNOR BIO 12/27-02/10 VIDAS LMX	LMX	37°C - 26/30h			ChromID 37°C - 24h
AFNOR BIO 12/11-03/04 VIDAS LMO2 (37°C)	Fraser 1/2	30°C - 24/26h	Fraser	37°C - 24/26h	Chromogenic medium / Palcam / Oxford
AFNOR AES 10/03-09/00 ALOA one day	Fraser 1/2	30°C - 24±2h			ALOA One Day 37°C - 24/48h
AFNOR BKR 23/02-11/02 Compass L. mono	Fraser 1/2	30°C - 24±2h			Compass Listeria Agar 37°C - 24h
AFNOR BRD 07/16-01/09 Agar Listeria	Fraser 1/2	30°C - 24±2h			Agar Listeria 37°C - 24h
AFNOR UNI 03/04-04/05 Listeria Precis	One Broth Listeria	30°C - 24±2h			Brilliance Listeria 37°C - 24h
AFNOR BIO 12/40-11/16 GENE UP LMO	LPT	35-37°C - 24±2h			ALOA 35-37°C - 24/48h
AFNOR BIO 12/18-03/06 VIDAS LDUO	LX	30°C - 24±2h	LX	30°C - 24/26h	Chromogenic medium / Palcam / Oxford
AFNOR UNI 03/08-11/13 PCR	LEB	37°C - 24/28h			Lysis + PCR

The detail of the methodology followed by 77 laboratories using NF ISO/EN ISO 11290-1 and NM ISO 11290-1 methods, and the 10 laboratories using another method, is clarified in the following table :

Parameter	Mode	Nb laboratories
Method	ISO/NF EN ISO 11290-1	54
	NM ISO 11290-1	23
	Other	10
Primary enrichment medium	None primary enrichment	1
	Half-Fraser	76
	One broth Listeria	3
	Other	7
Primary enrichment temperature	30°C	82
	37°C	4
Primary enrichment duration	22-27 h	85
	48 h	1
Secondary enrichment medium	None secondary enrichment	9
	Fraser	73
	Other	1
Secondary enrichment temperature	37°C	71
	30°C	2
Secondary enrichment duration	22-25 h	59
	48 h	14
Isolation medium	Palcam	56
	Ottaviani et Agosti	44
	Compass Listeria	30
	Oxford	11
	Rapid L'mono	4
	Brilliance Listeria	3
	Other	2
Isolation temperature	37°C	80
	30°C	3
Isolation duration	44-48 h	51
	24±1 h	32
Confirmation test	None	6
	Biochemical	54
	Biochemical + CAMP	20
	Other	4
Nb of colonies per plate	1	29
	2-3	7
	5	38
	10	1

3. ASSESSMENT OF PERFORMANCE (INDIVIDUEL REPORTS)

3.1. PERFORMANCES IN ENUMERATION

Performance is assessed on two criteria : **precision and trueness**.

The assigned value of the contamination used to assess the trueness and the reference standard deviation for the assessment of the precision are consensual values obtained with the results of all the participants. These values are obtained by robust estimation methods in order to eliminate influence of aberrant results. However some results are excluded of the statistical analysis. That is the case when laboratories do not give results for all contaminated units, when results are "less than CFU/g", when samples are analyzed after the deadline (time of receipt > 4 days after sending or time of analysis >15 days after sending) or when this information is not specified.

A statistical analysis has also be done to highlight potential relations between techniques used (delay of analysis, preservation temperature, preparation of the initial suspension, homogenization technique, resuscitation conditions , method used, media used, manufacturers of media, preparation mode, plating method, incubation conditions) and results obtained. We need to clarify that this statistical link is not involved in a cause - effect relationship. Indeed, this link may be due to a not documented factor.

When a significant statistical link is identified between use of a technique and the obtained results, the assessment of performance is done considering the influence of one or several factors involved if their effect translates into a contamination's difference higher than 0.15 log CFU/g for non-selective media or higher than 0.30 log CFU/g for selective media (these limits match with productivity limits of culture media usually recommended in the standard NF EN ISO 11133).

PRECISION

The precision reflects the repeatability (or reproducibility intra-laboratory) of your work.

The standard deviation of your results, s , is compared to the robust estimation of the standard deviation (reference standard deviation of precision), s^* , obtained with algorithm S from the standard ISO 13528 applied to all standard deviations obtained by laboratories included in the statistical analysis.

An index score is then calculated using the following formula : $i = (k-1) \cdot \frac{s^2}{s^{*2}}$ (with k, number of contaminated units and retained in the statistical analysis, usually 5).

The standard ISO 13528 do not provide warning and action limits for this score, so its interpretation is left to your discretion.

As an indicator, we suggest following values by analogy with those indicated for the evaluation of trueness.

For k=5, a score lower than 0.1 or higher than 18 may be considered as an action signal and a score lower than 0.45 or higher than 11.5 may be considered as a warning signal.

For k=4, a score lower than 0.03 or higher than 15.5 may be considered as an action signal and a score lower than 0.2 or higher than 9.5 may be considered as a warning signal.

For k=3, a score lower than 0.003 or higher than 13.2 may be considered as an action signal and a score lower than 0.05 or higher than 7.5 may be considered as a warning signal.

For k=2, a score lower than 0.000002 or higher than 10.3 may be considered as an action signal and a score lower than 0.0008 or higher than 5.2 may be considered as a warning signal.

TRUENESS

The trueness reflects the closeness of the mean of your results to the contamination's assigned value of samples. It has been evaluated for all enumerated flora.

The mean of your results in log CFU/g, m (on contaminated units and included in the statistical analysis), is compared to the contamination's assigned value, m_{pt} , obtained with algorithm A from the standard ISO 13528 applied to all laboratories mean included in the statistical analysis. When groups are formed, each one is characterized by its own assigned value.

A z score is then calculated with the following formula : $z = \frac{m - m_{pt}}{\sigma_{pt}}$, where σ_{pt} is the standard deviation for proficiency assessment (robust estimation of the standard deviation obtained by participants).

The standard ISO 13528 specifies that a z score lower than -3 or higher than +3 must be considered as an action signal and that a z score lower than -2 or higher than +2 must be considered as a warning signal.

In this report, we specify, estimations of interlaboratories standard deviation for enumerations proposed as well as reproducibility standard deviation or global standard deviation for the test (parameters including interlaboratories variability and the variability of the precision).

INDIVIDUAL REPORTS – FOR EACH CRITERIA YOU FIND THE FOLLOWING INFORMATIONS

- your results in logarithm base 10 (-1 when the answer is < limit and NaN when there is no answer).
Comment : the presentation order of your results does not necessarily correspond to the order you sent them, this order is the same for all the flora.
- histogram for the studied parameter (laboratories standard deviations for the precision and laboratory's means for the trueness) with an asterisk indicating the location of your result,
- standard deviation (precision) or mean (trueness) of your results (on contaminated units and retained in the statistical analysis),
- the method declared in your results input,
- when necessary, your group in relation to the technique used,
- precision score or z score,
- number of laboratories which made analysis (and belonging to your group),
- number of laboratories included in the statistical analysis,
- reference standard deviation for the precision or assigned value of the contamination and standard deviation aptitude assessment (trueness),
- number of laboratories with a satisfactory signal,
- number of laboratories with a warning signal,
- number of laboratories with an action signal.

3.1.1. MICROORGANISMS AT 30°C

None significant effect of the analysis technique has been highlighted.

Microorganisms at 30°C	
Assigned value of the contamination (log cfu/g)	4.870
Assigned value uncertainty (log cfu/g)	0.0062
Standard deviation for proficiency assessment (log cfu/g)	0.0845
Standard deviation for precision (log cfu/g)	0.0573
Interlaboratory's standard deviation (log cfu/g)	0.0805
Reproducibility standard deviation (log cfu/g)	0.0989

3.1.2. ENTEROBACTERIACEAE

A significant “effect” of the culture medium, manufacturer and the retained dilution has been highlighted. This effect results in a contamination’s difference higher than 0.3 log cfu/g, then results have been gathered in three groups :

Enterobacteriaceae	Group 1	Group 2	Group 3
Assigned value of the contamination (log cfu/g)	2.658	2.920	3.180
Assigned value uncertainty (log cfu/g)	0.0258	0.0663	0.0179
Standard deviation for proficiency assessment (log cfu/g)	0.2808	0.2652	0.0972
Standard deviation for precision (log cfu/g)		0.1168	
Interlaboratory's standard deviation (log cfu/g)	0.2759	0.2600	0.0820
Reproducibility standard deviation (log cfu/g)	0.2997	0.2850	0.1427

3.1.3. TOTAL COLIFORMS

A significant “effect” of the culture medium, manufacturer, and the preparation mode has been highlighted. This effect results in a contamination’s difference higher than 0.3 log cfu/g, then results have been gathered in three groups :

Total coliforms	Group 1	Group 2	Group 3
Assigned value of the contamination (log cfu/g)	2.416	2.658	3.160
Assigned value uncertainty (log cfu/g)	0.0403	0.0305	0.0368
Standard deviation for proficiency assessment (log cfu/g)	0.2659	0.2688	0.1350
Standard deviation for precision (log cfu/g)		0.1154	
Interlaboratory's standard deviation (log cfu/g)	0.2609	0.2638	0.1248
Reproducibility standard deviation (log cfu/g)	0.2858	0.2885	0.1709

3.1.4. THERMOTOLERANT COLIFORMS

A significant “effect” of the culture medium, manufacturer and the retained dilution has been highlighted. This effect results in a contamination’s difference higher than 0.3 log cfu/g, then results have been gathered in three groups :

Thermotolerant coliforms	Group 1	Group 2	Group 3
Assigned value of the contamination (log cfu/g)	2.469	2.626	2.912
Assigned value uncertainty (log cfu/g)	0.0252	0.0400	0.0924
Standard deviation for proficiency assessment (log cfu/g)	0.2101	0.2617	0.3388
Standard deviation for precision (log cfu/g)		0.1238	
Interlaboratory's standard deviation (log cfu/g)	0.2027	0.2558	0.3343
Reproducibility standard deviation (log cfu/g)	0.2339	0.2812	0.3541

3.1.5. *ESCHERICHIA COLI*

A significant “effect” of the preparation mode of the culture medium has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

<i>Escherichia coli</i>	
Assigned value of the contamination (log cfu/g)	2.188
Assigned value uncertainty (log cfu/g)	0.0173
Standard deviation for proficiency assessment (log cfu/g)	0.2247
Standard deviation for precision (log cfu/g)	0.1822
Interlaboratory's standard deviation (log cfu/g)	0.2094
Reproducibility standard deviation (log cfu/g)	0.2775

3.1.6. ANAEROBIC SULFITE-REDUCING BACTERIA

Only units n°1, 2 and 5 were artificially contaminated.

A significant “effect” of the manufacturer and the retained dilution has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

Anaerobic sulfite-reducing bacteria	
Assigned value of the contamination (log cfu/g)	2.764
Assigned value uncertainty (log cfu/g)	0.0200
Standard deviation for proficiency assessment (log cfu/g)	0.2379
Standard deviation for precision (log cfu/g)	0.0914
Interlaboratory's standard deviation (log cfu/g)	0.2319
Reproducibility standard deviation (log cfu/g)	0.2493

Comment :

- 5 laboratories detected ASR in unit n°3 non-artificially contaminated by *C. perfringens* with a contamination level from 20 to 1600 cfu/g.

- 5 laboratories detected ASR in unit n°4 non-artificially contaminated by *C. perfringens* with a contamination level from 10 to 2000 cfu/g.

3.1.7. CLOSTRIDIUM PERFRINGENS

Only units n°1, 2 and 5 were artificially contaminated.

A significant “effect” of the retained dilution has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

<i>Clostridium perfringens</i>	
Assigned value of the contamination (log cfu/g)	2.762
Assigned value uncertainty (log cfu/g)	0.0214
Standard deviation for proficiency assessment (log cfu/g)	0.2277
Standard deviation for precision (log cfu/g)	0.0929
Interlaboratory's standard deviation (log cfu/g)	0.2213
Reproducibility standard deviation (log cfu/g)	0.2400

Comment :

- 4 laboratories detected *C. perfringens* in unit n°3 non-artificially contaminated by *C. perfringens* with a contamination level from 400 to 1600 cfu/g.
- 1 laboratory detected *C. perfringens* in unit n°4 non-artificially contaminated by *C. perfringens* with a contamination level of 1000 cfu/g.

3.1.8. COAGULASE POSITIVE STAPHYLOCOCCI

A significant “effect” of the resuscitation’s duration has been highlighted. This effect results in a contamination’s difference lower than 0.3 log cfu/g, then results have been gathered in one group :

Coagulase positive Staphylococci	
Assigned value of the contamination (log cfu/g)	3.612
Assigned value uncertainty (log cfu/g)	0.0112
Standard deviation for proficiency assessment (log cfu/g)	0.1490
Standard deviation for precision (log cfu/g)	0.0802
Interlaboratory's standard deviation (log cfu/g)	0.1446
Reproducibility standard deviation (log cfu/g)	0.1654

3.1.9. LISTERIA MONOCYTOGENES

Only units n°1, 2 and 5 were artificially contaminated.

None significant effect of the analysis technique has been highlighted.

<i>Listeria monocytogenes</i>	
Assigned value of the contamination (log cfu/g)	3.408
Assigned value uncertainty (log cfu/g)	0.0086
Standard deviation for proficiency assessment (log cfu/g)	0.1040
Standard deviation for precision (log cfu/g)	0.0732
Interlaboratory's standard deviation (log cfu/g)	0.0950
Reproducibility standard deviation (log cfu/g)	0.1200

3.2.PERFORMANCES IN DETECTION

The performance is assessed by the capacity to detect only samples contaminated by *Salmonella* and *Listeria monocytogenes* (no false positive or false negative results).

3.2.1. DETECTION – *SALMONELLA*

Only units n°1 and 5 were artificially contaminated.

294 laboratories obtained correct results.

7 laboratories obtained false positive results (respectively 4, 2 and 3 false-positive for units n° 2, 3 and 4).

6 laboratories obtained false negative results (respectively 3 and 4 false-negative for units n° 1 and 5).

3.2.2. DETECTION – *LISTERIA MONOCYTOGENES*

Only units n°1, 2 and 5 were artificially contaminated.

266 laboratories obtained correct results.

2 laboratories obtained false positive results (respectively 1 and 1 false-positive for units n°3 and 4).

2 laboratories obtained false negative results (respectively 1, 1 and 2 false-negative for units n°1, 2 and 5).

3.3.EVOLUTION OF PERFORMANCE

You will find, on each page of your performance's assessment, a graph representing evolution of it on different tests since the 53th scheme.

In order to interpret your control card with z scores, you can refer to the standard ISO 13528 §10.8.2.2, explaining the 3 « out of control » situations:

- Just one overtaking of the action limit ($z < -3$ or $z > 3$),
- 2 consecutives z scores out of 3 overtaking of the warning limit ($2 < z < 3$ or $-3 < z < -2$),
- 6 consecutives z scores regularly increasing or decreasing.